

Manual Supplement

Manual Title:	Metrology Software Install & Upgrade Guide Version 6		
Part Number:	1275389	Supplement Issue:	2
Print Date:	July 1999	Issue Date:	9/01
Revision/Date:		Page Count:	2

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1

On pages 4-18 through 4-20, replace the entire **Item 6: Update the Database** section with the following:

Item 6: Update the Database

The following procedure applies to users performing an upgrade installation on a network from version 4.0 or later.

Note

The DBUPDATE.EXE application must be executed from the server or from a stand-alone workstation.

The application must have access to Sybase SQL Anywhere Server applications as well as the actual database files. If the server computer does not have any client applications installed, use the client installation setup program which exists in the CLIENT sub-folder of the MET/BASE shared executable directory.

1. Run SETUP.EXE from this directory. You may install any client application. This will set up the ODBC data source on the server and enable the DBUPDATE.EXE application to login to the database.
2. Ensure that the Calibration Data Engine has been started.
3. Execute DBUPDATE.EXE.
4. Follow DBUPDATE instructions carefully. Many of the instructions on the screen require specific user actions or the update may abort.
5. Locate and run the program DBUPDATE.EXE.
6. Login is as the MT user, with the password MT, or with your password if you are upgrading an existing system.
7. Follow the on-screen instructions as they appear.

WARNING

You MUST stop the Calibration Data Engine when you are prompted by the system. Failure to stop the Calibration Data Engine when you are prompted to do so completely interrupts the updating of the database, with unpredictable results and possible loss of data.

8. If there are more workstations to update, return to “Item 5: Install a Client” to continue.
9. If you desire, the Fluke Metrology Software – Client Uninstall program may be run from the system Control Panel to remove the client functions from the server machine without affecting the operation of the server.
10. If this is the last (or only) workstation you are upgrading, check off *Item* ✓6 on the flowchart at the beginning of this chapter, and continue to the next item to apply the licenses.

Change #2

On the back of the front cover change the FAX #,

From: 1-425-356-5992

To: 1-425-446-5992

On page 1-2, change the following diagram,

From:

I have MET/TRACK Version 4.x or 5.x MET/CAL 4.x or 5.x Database Mode

To:

I have MET/TRACK Version 4.x or later MET/CAL 4.x or later Database Mode

On page 1-3, under ***Supported Platforms***, replace the last sentence with the following:

However, the main text of this manual describes only installation on machines running Microsoft Windows 95, Windows 98, Windows ME, Windows NT, Windows 2000 or Windows XP.

On page 1-5, under **Other Documentation**, add the following sentence to the end of the first paragraph:

This reader installation AR32E01.EXE and the PDF files will be installed in the shared MET/CAL directory on the Server so that the manuals may be accessed from a Standalone or Client installation without requiring the installation disk.

On page 1-7, under *Note*, delete the second, third and fourth sentence.

Under **Process Overview**, replace the last sentence on the page with the following:

The Crystal Reports software should be loaded on a client or standalone workstation.

On page 2-3, under *Note*, delete the second, third and fourth sentences.

On page 3-3, under *Note*, delete the second, third and fourth sentences.

On page 4-2, under **Network Upgrade Flowchart**, replace the title of the chart with the following:

**FLOWCHART 4 UPGRADING A NETWORKED DATABASE MODE
MET/CAL 4.2, MET/TRACK 4.X, MET/BASE 5.x OR MET/BASE 6.x**

Under Item 5, add the following Note prior to step 1:

Note: Steps 1 and 2 apply to Version 4.x and 5.x only.

On page 4-3, under **MET/CAL Users**, replace the first sentence with the following:

The instructions in this chapter describe how to upgrade from Version 4.2 or later in Database Mode Version 6 Database Mode.

Under *Note* number 1, delete the second, third and fourth sentences.

On page 4-4, replace the title of Step E with the following:

E: For Upgrades from Version 4.x and 5.x Only, Delete WATCOM, MET/TRACK and MET/CAL Icons and Shortcuts

On page 5-3, under **MET/CAL Users**, replace the first sentence with the following:

The instructions in the chapter describe how to upgrade from Version 4.2 or later in Database Mode to Version 6 Database Mode.

Under *Note*, delete the second, third and fourth sentences.

On page 5-4, under, Step D, in the first paragraph,

Change: satabase

To: database

On page 6-2, replace the entire entry under Item 1A with the following:

Obtain Migration from MET/SUPPORT. These files are no longer provided on MET/BASE CD-ROM.

On page 6-5, under Item 1A, replace the first sentence with the following:

In performing the first item, you retrieve the migration programs from MET/SUPPORT, copying them onto the local hard drive.

On page 8-5, under **Installing Acrobat Reader**, replace the first sentence with the following:

In order to install and use electronic manuals provided on your MET/CAL CD, or from the shared MET/CAL file folder, you must have Acrobat Reader installed on your computer.

Replace Step 1 with the following:

1. Use the Windows Explorer to view the contents of MET/CAL CD or the shared MET/CAL file folder.

On page 8-6, replace Step 1 with the following:

1. Use the Windows Explorer to view the contents of MET/CAL CD or the shared MET/CAL file folder.

Under Step 3, add the following:

Note

It may be helpful to set up Windows shortcuts to the PDF manual files in the Fluke Metrology Software shortcut folder on the START Menu.

On page A-1, under **Mail, FAX, E-mail and World Wide Web**, change the prefix on the Phone and the Fax numbers:

From: 356

To: 446

On page B-1 and B-2, replace the entire Appendix B with the following:

The facility for creating a 3.5 floppy disk installation set is no longer provided on the MET/BASE CD.

