



IVI-COM Instrument Driver Programming Guide (Excel2000 VBA Edition)

Dec 2003 Revision 1.0

1- Overview

1-1 Using IVI-COM Drivers in Excel 2000 VBA

Excel VBA is one of the most suitable development environments for use with IVI-COM instrument drivers. Since COM programming style such as using ActiveX controls is very popular in Excel VBA, many Excel VBA programmers are familiar with using them. Although an IVI-COM instrument driver is not an ActiveX control, you can develop your programs in the same manner that when you use generic COM objects.

When using an IVI-COM instrument driver, there are two approaches – using specific interfaces and using class interfaces. The former is to use interfaces that are specific to an instrument driver and you can utilise the most of features of the instrument you use. The later is to utilise instrument class interfaces that are defined in the IVI specifications allowing to utilise interchangeability features, but instrument specific features are restricted.

Notes:

The instrument class to which the instrument driver belongs is documented in Readme.txt for each of drivers. The Readme document can be viewed from Start button→Program→IVI folder.

If the instrument driver does not belong to any instrument classes, you can't utilise class interfaces. This means that you cannot develop applications that utilise interchangeability features.

This guidebook assumes that you use Excel 2000. Operations on other Excel versions may be slightly different.

1-2 Creating An Application Project

This document explains how to develop a form-oriented application using Excel VBA. After launching Excel, select **Tools | Macro | Visual Basic Editor** menu to start up the VBA integrated development environment, then add a user form by selecting **Insert | UserForm** menu.

Notes:

This guidebook assumes that you use IVI-COM Kikusui4800 instrument driver (for KIKUSUI PIA4800 series DC Power Supply Controller). You can also use IVI-COM instrument drivers for other models in the same manner.

2- Example Using Specific Interfaces

Here we introduce an example using specific interfaces. By using specific interfaces, you can utilise the maximum power of driver features but you have to spoil interchangeability.

2-1 Importing Type Libraries

What you should do first after creating a new project is import the type libraries of IVI-COM instrument drivers you want to use. Choose **Tools | References** menu to bring up the **References** dialogue. Since this example assumes that you use Kikusui4800 IVI-COM driver, you need check **Kikusui4800 (Kikusui) 1.0 Type Library**.




Figure 2-1 Importing Type Libraries

Importing type libraries is now complete so your application can use the specific interfaces provided by the Kikusui4800 IVI-COM driver.

2-2 Object Browser

Once the type libraries have been imported, you can confirm COM interface syntaxes through the Excel VBA Object Browser. To launch the Object Browser, choose **View | Object Browser** menu (or press **F2** key).




Figure 2-2 Object Browser

2-3 Creating Object and Initialising Session

First, doubleclick on the design-time form with the mouse. Then the `UserForm_Click` event handler having only a skeleton code will be shown, however, select the `UserForm_Initialize` event since the code should be written in it. Furthermore, declare `m_dcpwr` as a form's data member variable as `Kikusui4800Lib.Kikusui4800` type. Do not forget to write the `New` operator since you also create an instrument driver object here.




Figure 2-3 UserForm_Initialize handler

Just creating the object does not perform any instrument I/Os. To initiate I/Os with the instrument, you use the `Initialize` method. As you type `m_dcpwr.` in the `UserForm_Initialize` handler, the IntelliSense feature of Excel VBA will show the method/property list for `Kikusui4800` type. Here you select `Initialize` method then press the Tab key.




Figure 2-4 IntelliSense (Method/Property List)

Furthermore, by typing a space or a left parenthesis "(" after `Initialize`, IntelliSense will show all the parameters for the method.




Figure 2-5 IntelliSense (Parameter List)

Now let's talk about the parameters for the `Initialize` method. Every IVI-COM instrument driver has an `Initialize` method that is defined in the IVI specifications. This method has the following parameters.

Table 2-1 Parameters for Initialize method

Parameter	Type	Description
ResourceName	String	VISA resource name string. This is decided according to the I/O interface and/or address through which the instrument is connected. If the instrument has the address 3 on the GPIB board #0, for example, it can be GPIB0::3::INSTR.
IdQuery	Boolean	Specifying TRUE performs ID query to the instrument.
Reset	Boolean	Specifying TRUE resets the instrument settings.
OptionString	String	Overrides the following settings instead of default: RangeCheck Cache Simulate QueryInstrStatus RecordCoercions Interchange Check Furthermore you can specify driver-specific options if the driver supports DriverSetup features.

ResourceName specifies a VISA resource. If IdQuery is TRUE, the driver queries the instrument identities using a query command such as "*IDN?". If Reset is TRUE, the driver resets the instrument settings using a reset command such as "*RST".

OptionString has two features. One is what configures IVI-defined behaviours such as RangeCheck, Cache, Simulate, QueryInstrStatus, RecordCoercions, and Interchange Check. Another one is what specifies DriverSetup that may be differently defined by each of instrument drivers. Because the OptionString is a string parameter, these settings must be written as like the following example:

```
QueryInstrStatus = TRUE , Cache = TRUE , DriverSetup=12345
```

Names and setting values for the features being set are case-insensitive. Since the setting values are Boolean type, you can use any of TRUE, FALSE, 1, and 0. Use commas for splitting multiple items. If an item is not explicitly specified in the OptionString parameter, the IVI-defined default value is applied for the item. The IVI-defined default values are TRUE for RangeCheck and Cache, and FALSE for others.

Some instrument drivers may have special meanings for the DriverSetup parameter. It can specify items that are not defined by the IVI specifications when invoking the Initialize method, and its purpose and syntax are driver-specific. Therefore, specifying the DriverSetup must be at the last part on the OptionString parameter. Because the contents of DriverSetup are different depending on each driver, refer to driver's Readme document or online help.

Now try to write Initialize call. The OptionString parameter is optional and you can skip it.

```
Private Sub UserForm_Initialize()  
    m_dcpwr.Initialize "GPIB0::3::INSTR", True, True  
End Sub
```


2-4 Closing Session

To close the instrument driver session, use the `Close` method. Since this example wrote the `Initialize` method call in the `UserForm_Initialize` handler, it is better to write `Close` method call in the `UserForm_Terminate` handler.

```
Option Explicit
Private m_dcpwr As New Kikusui4800Lib.Kikusui4800

Private Sub UserForm_Initialize()
    m_dcpwr.Initialize "GPIB0::3::INSTR", True, True
End Sub

Private Sub UserForm_Terminate()
    m_dcpwr.Close
End Sub
```

2-5 Execution

You can execute the previous codes for the time being. Since the `Initialize` method is invoked in the `UserForm_Initialize` handler, communications with the instrument immediately start as you launch the program. If the instrument is actually connected and the `Initialize` method call has succeeded, the form screen appears. If a communication problem has occurred or the VISA library is not configured properly, a COM exception (VBA runtime error) will be generated.




Figure 2-6 COM exception

2-6 Repeated Capabilities

In case of Kikusui4800 IVI-COM driver, output settings for the DC power supplies are performed through the Output interfaces as the same concept defined by the `IviDCPwr` class. In case of specific interfaces provided by the Kikusui4800 driver, they are the `IKikusui4800Output` and `IKikusui4800Outputs` interfaces. An instrument driver that is compliant with the `IviDCPwr` class is designed assuming that the instrument is a multi-track power supply equipping multiple output channels.

These COM interfaces have the same name with an exception of differences between singular and plural forms. An interface having this kind of plural name is generally called "repeated capabilities" in the IVI specifications. Repeated capabilities are something like a container that is defined for handling equivalent or similar multiple objects, and a COM interface having a plural name such as `IKikusui4800Outputs` normally has the `Count`, `Name`, and `Item` properties (all are read-only). Plus, a singular object can be referenced through the `Item` property.

First look at the following example, which controls an output channel identified by the name "N5!C1" on the power supply instrument (actually a Kikusui PIA4800 series DC Power Supply Controller) hosted by the Kikusui4800 IVI-COM driver. This example writes the codes in the event handler assuming you put a command button (`CommandButton1`) on the form.


```
Private Sub CommandButton1_Click()  
    Dim out As Kikusui4800Lib.IKikusui4800Output  
    Set out = m_dcpwr.Outputs.Item("N5!C1")  
  
    out.VoltageLevel = 10.5  
    out.CurrentLimit = 1.2  
    out.Enabled = True  
End Sub
```

Once the `IKikusui4800Output` interface has been acquired, there is no difficulty at all. The `VoltageLevel` and the `CurrentLimit` properties set voltage level and current limit settings respectively. The `Enabled` property switches output ON/OFF state.

Mind the grammar for acquiring the `IKikusui4800Output` interface. This example here acquires the `IKikusui4800Outputs` interface though the `Output` property of the `IKikusui4800` interface, then acquires `IKikusui4800Output` interface by using the `Item` property.

```
Dim out As Kikusui4800Lib.IKikusui4800Output  
Set out = m_dcpwr.Outputs.Item("N5!C1")
```

The codes can also be written as like below.

```
Dim outs As Kikusui4800Lib.IKikusui4800Outputs  
Dim out As Kikusui4800Lib.IKikusui4800Output  
Set outs = m_dcpwr.Outputs  
Set out = outs.Item("N5!C1")
```

Now mind the parameter passed to the `Item` property. This parameter specifies the name of the single `Output` object to be referenced. Actual available names (`Output Name`) are however different depending on drivers. For example, Kikusui4800 IVI-COM driver uses an expression like "N1!C1" specifying `NODE` and `CH`. However other drivers, even if being `IviDCPwr` class-compliant, may have different names. One instrument driver, for example, may use an expression like "Track1". Although available names on a particular instrument driver are normally documented in the driver's online help, you can also check them out by writing some test codes shown below.

```
Dim outs As Kikusui4800Lib.IKikusui4800Outputs  
Set outs = m_dcpwr.Outputs  
  
Dim cnt As Long  
Dim ndx As Long  
cnt = outs.Count  
For ndx = 1 To cnt  
    Dim strName As String  
    strName = outs.Name(ndx)  
    Debug.Print strName  
Next ndx
```

The `Count` property returns number of single objects that the repeated capabilities have. The `Name` property returns the name of single object for the given index. The name is exactly the one that can be passed to the `Item` property as a parameter. In the above example, the codes iterate from the index 1 to `Count` by using the `For/Next` statement. Mind that the index numbers for the `Name` parameter is one-based, not zero-based.

3- Example Using Class Interfaces

Now we explain how to use class interfaces. By using class interfaces, you can swap the instruments without recompiling/relinking your application codes. In this case, however, IVI-COM instrument drivers for both pre-swap and post-swap models must be provided, and these drivers both must belong to the same instrument class. There is no interchangeability available between different instrument classes.

3-1 Virtual Instrument

What you have to do before creating an application that utilises interchangeability features is create a virtual instrument. To realise interchangeability features, you should not write codes that are very specific to a particular IVI-COM instrument driver (e.g. creating an object instance directly as Kikusui4800 type) and should not write a specific VISA resource name such as "GPIB0::3::INSTR". Writing them directly in the application spoils interchangeability.

Instead, the IVI-COM specifications define methods to realise interchangeability by placing an external IVI configuration store. The application indirectly selects an instrument driver according to contents of the IVI Configuration Store, and accesses the indirectly loaded driver through the class interfaces.

The IVI Configuration Store is normally /Program Files/IVI/Data/IviConfigurationStore.XML file and is accessed through the IVI Configuration Server DLL. This DLL is mainly used by IVI-COM instrument drivers and some configuration tools provided by instrument driver vendors, not by end-user applications. KIKUSUI provides a configuration tool called **Kikusui IVI Config Utility** that allows you to configure virtual instrument settings.

Notes:

As for how to configure virtual instruments by using Kikusui IVI Config Utility, refer to "Programming Guide, (IVI Config Utility Edition)."

This guidebook assumes that a virtual instrument having the logical name "MySupply" is already created, using Kikusui4800 driver, and using a VISA resource "GPIB0::3::INSTR".

3-2 Importing Type Libraries

What you should do first after creating a new project is import the type library of IVI-COM class interfaces that you want to use. Choose **Tools | References** menu to bring up the **References** dialogue. Since we use IviDCPwr class interfaces, check **IviDCPwr 2.0 TypeLibrary**. Furthermore, make sure to check **IviDriver 1.0 Type Library** and **IviSessionFactory 1.0 TypeLibrary** regardless instrument classes you use.




Figure 3-1 Importing Type Libraries

Importing type libraries are now completed. Your application will be able to use arbitrary instrument drivers through the IviDCPwr class interfaces.

3-3 Object Browser

Once the type libraries have been imported, you can confirm COM interface syntaxes through the Visual Basic 6.0 Object Browser. To launch the Object Browser, choose **View | Object Browser** menu (or press **F2** key).




Figure 3-2 Object Browser

3-4 Creating Object and Initialising Session

First, doubleclick on the design-time form with the mouse. Then the UserForm_Initialize event handler having only a skeleton code will be shown. Declare m_dcpwr as a form's data member variable as IviDCPwrLib.IIviDCPwr type. A type

that begins with capital "I" such as IIVI DCPwr is in general a COM interface, therefore you cannot create an object with the New operator.




Figure 3-3 UserForm_Initialize handler

Next you must create an instrument driver object. To create an instrument driver object, use IVI Session Factory. The IVI Session Factory is a DLL server that comes with the IVI Shared Components. It extracts configuration information of the virtual instrument specified by the given logical name, loads the appropriate instrument driver software, then creates an instrument driver object.

```
Private Sub UserForm_Initialize()  
  
    Dim sf As New IVISESSIONFACTORYLib.IviSessionFactory  
    Set m_dcpwr = sf.CreateDriver("MySupply")  
  
    m_dcpwr.Initialize "MySupply", True, True  
  
End Sub
```

The example here creates an IviSessionFactory object and then creates an instrument object with the CreateDriver method. The parameter "MySupply" must already be configured as a valid logical name. The instrument driver DLL to be loaded will be what specified by the virtual instrument MySupply.

After creating the driver object, invoke the Initialize method. Although parameters of the Initialize method are exactly the same as the case of using specific interfaces, you should specify the logical name for the ResourceName parameter instead of specifying a VISA resource.

Again let's talk about parameters of the Initialize method. Every IVI-COM instrument driver has an Initialize method that is defined by the IVI specifications. This method has the following parameters.

Table 3-1 Parameters for Initialize method

Parameter	Type	Description
ResourceName	String	VISA resource name string. This is decided according to the I/O interface and/or address through which the instrument is connected. If the instrument has the address 3 on the GPIB board #0, for example, it can be GPIB0::3::INSTR. If specifying a logical name, the VISA resource that is described in the logical name's Hardware Asset configuration will be indirectly specified.
IdQuery	Boolean	Specifying TRUE performs ID query to the instrument.
Reset	Boolean	Specifying TRUE resets the instrument settings.
OptionString	String	Overrides the following settings instead of default: RangeCheck Cache Simulate QueryInstrStatus RecordCoercions Interchange Check Furthermore you can specify driver-specific options if the driver supports DriverSetup features.

In general, applications that are aware of interchangeability use a logical name for the ResourceName parameter rather than a VISA resource. Actually it is possible to specify a VISA resource, however, doing so slightly spoils abstraction of virtual instrument.

If IdQuery is TRUE, the driver queries the instrument identities using a query command such as "*IDN?". If Reset is TRUE, the driver resets the instrument settings using a reset command such as "*RST".

OptionString has two features. One is what configures IVI-defined behaviours such as RangeCheck, Cache, Simulate, QueryInstrStatus, RecordCoercions, and Interchange Check. Another one is what specifies DriverSetup that may be differently defined by each of instrument drivers. Because the OptionString is a string parameter, these settings must be written as like the following example:

```
QueryInstrStatus = TRUE , Cache = TRUE , DriverSetup=12345
```

Names and setting values for the features being set are case-insensitive. Since the setting values are Boolean type, you can use any of TRUE, FALSE, 1, and 0. Use commas for splitting multiple items. If an item is not explicitly specified in the OptionString parameter, the IVI-defined default value is applied for the item. The IVI-defined default values are TRUE for RangeCheck and Cache, and FALSE for others.

Some instrument drivers may have special meanings for the DriverSetup parameter. It can specify items that are not defined by the IVI specifications when invoking the Initialize method, and its purpose and syntax are driver-specific. Therefore, specifying the DriverSetup must be at the last part on the OptionString parameter. Because the contents of DriverSetup are different depending on each driver, refer to driver's Readme document or online help.

Now try to write Initialize call. The OptionString parameter is optional and you can skip it.


```
m_dcpwr.Initialize "MySupply", True, True
```

3-5 Closing Session

To close the instrument driver session, use the `Close` method. Since this example wrote the `Initialize` method call in the `UserForm_Initialize` handler, it is better to write `Close` method call in the `UserForm_Terminate` handler.

```
Option Explicit
Private m_dcpwr As IviDCPwrLib.IIviDCPwr

Private Sub UserForm_Initialize()

    Dim sf As New IVISESSIONFACTORYLib.IviSessionFactory
    Set m_dcpwr = sf.CreateDriver("MySupply")

    m_dcpwr.Initialize "MySupply", True, True

End Sub

Private Sub UserForm_Terminate()
    m_dcpwr.Close
End Sub
```

3-6 Execution

You can execute the previous codes for the time being. Since the `Initialize` method is invoked in the `UserForm_Initialize` handler, communications with the instrument immediately start as you launch the program. If the instrument is actually connected and the `Initialize` method call has succeeded, the form screen appears. If a communication problem has occurred or the VISA library is not configured properly, a COM exception (Visual Basic 6.0 runtime error) will be generated.




Figure 3-4 COM exception

3-7 Repeated Capabilities

In case of `IviDCPwr` class interfaces, output settings for the DC power supplies are performed through the Output interfaces. The interfaces used here are `IIviDCPwrOutput` and `IIviDCPwrOutputs`. An instrument driver that is compliant with the `IviDCPwr` class is designed assuming that the instrument is a multi-track power supply equipping multiple output channels.

These COM interfaces have the same name with an exception of differences between singular and plural forms. An interface having this kind of plural name is generally called "repeated capabilities" in the IVI specifications. Repeated capabilities are something like a container that is defined for handling equivalent or similar multiple objects, and a COM interface having a plural name such as `IIviDCPwrOutputs` normally has the `Count`, `Name`,

and Item properties (all are read-only). Plus, a singular object can be referenced through the Item property.

First look at the following example, which controls an output channel identified by the name "Track_A" registered as in the virtual instrument's virtual name. This example writes the codes in the event handler assuming you put a command button (CommandButton1) on the form.

```
Private Sub CommandButton1_Click()  
    Dim out As IviDCPwrLib.IIviDCPwrOutput  
    Set out = m_dcpwr.Outputs.Item("Track_A")  
  
    out.VoltageLevel = 10.5  
    out.CurrentLimit = 1.2  
    out.Enabled = True  
End Sub
```

Once the IIviDCPwrOutput interface has been acquired, there is no difficulty at all. The VoltageLevel and the CurrentLimit properties set voltage level and current limit settings respectively. The Enabled property switches output ON/OFF state.

Mind the grammar for acquiring the IIviDCPwrOutput interface. This example here acquires the IIviDCPwrOutputs interface though the Output property of the IIviDCPwr interface, then acquires IIviDCPwrOutput interface by using the Item property.

```
Dim out As IviDCPwrLib.IIviDCPwrOutput  
Set out = m_dcpwr.Outputs.Item("Track_A")
```

The codes can also be written as like below.

```
Dim outs As IviDCPwrLib.IIviDCPwrOutputs  
Dim out As IviDCPwrLib.IIviDCPwrOutput  
Set outs = m_dcpwr.Outputs  
Set out = outs.Item("Track_A")
```

Now mind the parameter passed to the Item property. This parameter specifies the name of the single Output object to be referenced. In the example that used specific interfaces, we specified a name that was very dependent upon each driver (physical name), however we use a different approach. Because a physical name that depends on a particular instrument driver should not be used, we specify a virtual name instead.

3-8 Swapping Instruments

In the previous examples, we used the Kikusui4800 instrument driver for the virtual instrument configurations. Now what will happen if you replace the instrument with the one that is hosted by the AgilentE36xx driver? In this case you do not have to recompile/relink your application, but you need change the virtual instrument configurations.

The configurations you have to change are basically Software Module selection on the Driver Session tab, and virtual name mappings on the Virtual Names tab (because physical names of the map target are changed). Replacing instruments may not allow using the same I/O interface (such as changing from a GPIB-only instrument to an RS232-only instrument), so you may have to change IO Resource Descriptor on the Hardware Asset tab as need.

Notes:

As for how to configure virtual instruments by using Kikusui IVI Config Utility, refer to "Programming Guide, (IVI Config Utility Edition)."

4- Error Handling

In the previous examples, there was no error handling processed. However, setting an out-of-range value to a property or invoking an unsupported function may generate an error from the instrument driver. Furthermore, no matter how the application is designed and implemented robustly, it is impossible to avoid instrument I/O communication errors.

When using IVI-COM instrument drivers, every error generated in the instrument driver is transmitted to the client program as a COM exception. In case of Excel VBA, a COM exception can be handled by using `On Error Goto` statement.

Now let's change the example of setting voltage and current as follows.

```
Private Sub CommandButton1_Click()  
  
    On Error GoTo DRIVER_ERR:  
  
    Dim out As Kikusui4800Lib.IKikusui4800Output  
    Set out = m_dcpwr.Outputs.Item("N5!C1")  
  
    out.VoltageLevel = 10.5  
    out.CurrentLimit = 1.2  
    out.Enabled = True  
  
    Exit Sub  
DRIVER_ERR:  
    MsgBox Err.Description, vbOKOnly, "Err 0x" & Hex(Err.Number)  
  
End Sub
```

In this example, errors are handled by using `On Error Goto` statement. For example, if the name passed to the `Item` property is wrong, if an out-of-range value is passed to `VoltageLevel`, or if an instrument communication error is generated, a COM exception will be generated in the instrument driver. Above example just displays a simple message box when an exception has occurred.

Detail about the error (COM exception) can be acquired through the `Err` object, which is defined in the Excel VBA. This example sets the error code (hexadecimal) obtained from `Number` property to the message box caption, and sets the description string obtained from the `Description` property to the main body text.




Figure 4-1 Message box by error handling

IVI-COM Instrument Driver Programming Guide

Product names and company names that appear in this guidebook are trademarks or registered trademarks of their respective companies.

©2003 Kikusui Electronics Corp. All Rights Reserved.