

DECLARATION OF CONFORMITY

according to ISO/IEC Guide 22 and EN 45014:1998

Manufacturer's Name: LeCroy Corporation
Manufacturer's Address: 700 Chestnut Ridge Road
Chestnut Ridge, NY 10977
USA

herewith declare that

Product(s) Name: Active Voltage Probe
Model Number(s): HFP2500, HFP1500, HFP1000

are in conformity with the provisions of the following EC directive(s), including the latest amendments, and with national legislation implementing these directives:

73/23/EEC Low Voltage Directive
89/336/EEC EMC Directive

and that conformity with Council Directive 73/23/EEC is based on

EN 61010-1: 1993+A2:1995 Safety requirements for electrical equipment for measurement control and laboratory use
Part 1: General requirements

EN 61010-2-031: 1994 Part 2-031: Particular requirements for hand-held probe assemblies for electrical measurement and test

and that conformity with Council Directive 89/336/EEC is based on

EN 61326-1: 1997+A1: 1998 EMC requirements for electrical equipment for measurement control and laboratory use

Emissions: EN 55011:1998 Class A Radiated Emissions

Immunity: EN61000-4-2: 1995 Electrostatic Discharge
(Air/Contact Discharge: 4 kV)

EN 61000-4-3: 1996 RF radiated Electromagnetic Field
(80 MHz..... 1 GHz; 3 V/m)



By: R. Kalyandrug
Product Compliance manager

Place: LeCroy Corporation
700 Chestnut Ridge Road
Chestnut Ridge, NY 10977
USA

Date: November 22, 2000

European Contact:
Your local LeCroy Sales Office or

LeCroy Europe GmbH
Waldhofer Str. 104
D-69123 Heidelberg
Germany

Tel: (49) 6221 82700
Fax: (49) 6221 834655

Warning: This is a Class A product. In a domestic environment this product may cause radio interference, in which case the user may be required to take adequate measures.

BLANK PAGE