

# Megohmmeter

# Model 24508

Code: 24508 EN

Delivery: ex stock

Warranty: 24 months



- Resistance measurement range from 50 x 10<sup>3</sup> ... 10 x 10<sup>12</sup>
- Current measurement range 10 x 10<sup>-12</sup> ... 10 x 10<sup>-3</sup> A
- Automatic / manual switch of measurement range
- Test voltage 45 V, 100 V, 250 V, 500 V
- Limit value indicator
- RS232 interface

### **Application**

Based on its specifications, this device can be used in various applications. It is especially suitable for resistance measurement on insulating materials such as e.g. cable insulations, foils, textiles, surfaces, insulating liquids, etc. With a test voltage of 45 V, 100 V, 250 V and 500 V the device fulfils most test specifications such as e.g. DIN 51953, 53482 and 54345.

The guard switching allows single resistance measurements in a triangle wiring. This could be e.g. a two line cable with common shield or the measurement of insulating materials on a guardring measurement cell.

The selection of the measurement range is done manually or automatically. Fast subsequent measurements can be realized by the internal limit value indicator. When the measured value exceeds the limit the limit value indicator switches and activates a potential-free relay output. The megohmmeter is the right instrument for use in laboratory as well as industrial applications.

### **Description**

The digital megohmmeter model 24508 is a microprocessorcontrolled measurement device for insulation resistances. The device has an easy-to-use structure in a sturdy metal housing. Easy access to the interior components allows an optimal service.

The measurement range stretches from  $50~\text{k}\Omega$  up to  $10~\text{T}\Omega$  resp. 10 pA up to 10 mA with a test voltage of 45 V, 100 V, 250 V and 500 V. The configuration of the device is done via the two line LCD display with the help of the simple menu structure. It goes without saying that all configurations can also be effected the RS232 interface. The connections for the potential-free limit output as well as the external measurement start / stop are located on the backside.

#### **Technical Data**

 $50 \text{ k}\Omega \dots 10 \text{ T}\Omega$ Resistance measurement range:

divided in 8 measurement ranges

 $50 \text{ k}\Omega$  ...  $1 \text{ T}\Omega$ :  $2.5 \% \text{ rdg.} \pm 1 \text{ Digit}$ Measurement accuracy:

1 T $\Omega$  ... 10 T $\Omega$ : 10 % rdg.  $\pm$  1 Digit

Current range: 10 pA ... 10 mA divided in 8 measurement ranges

Measurement accuracy: 500 pA ... 10 mA: 2.5 % rdg. ± 1 Digit

10 pA ... 500 pA: 10 % rdg. ± 1 Digit

Measurement voltage: 45 V, 100 V, 250 V, 500 V (other voltages upon request)

Measurement time: freely selectable up to 999 s

Max. current in measurement circuit:

Measurement range selection: manual or automatic Measurement connections: BNC (red) measurement voltage

> BNC (black) measurement input 4 mm ø socket (blue) guard

4 mm ø socket (green) ground

Display: two line LCD display measurement value 3 digits with unit

Limit value indicator: potential-free relay output

(max. 48 V, 1 A)

External measurement start: via potential-free contact

RS232 with 9 pin Sub Min D socket Interface:

0 °C ... 45 °C Operating temperature range:

Storage temperature: - 20 °C ... + 70 °C

Supply voltage: 230 V ± 10 % 50 Hz acc. to standard EN 61010-1

Device security: < 10 VA Power:

metal housing Housing:

Dimensions (W x H x D): 255 x 125 x 270 [mm]

Net weight:

## **Order Information**

**Digital Megohmmeter** Model 24508

incl. measurement leads 1 m length

and RS232 cable

## **Accessories**

Measurement leads 3 m length Model 24508-Z001 6-pin plug for ext. start and limit output **Model 9942** Model 24DKD-24508 **DKD Calibration Certificate** WKS Calibration Certificate Model 24WKS-24508

Guard ring electrodes for the measurements

of surface or volume resistances on request

# Calibration resistances for the device testing

Model series 1270



Operating voltage: 20 V ... 1000 V Temperature coefficient: ± 0.15 %/K typically maximum  $\pm$  0.30 %/K

Construction: metal housing with PVC cover

Dimensions: 36 x 30 x 90 [mm]

Net weight: approx. 70 g

Model	Resistance Value		Accuracy		Voltage Coefficient		
1270	10 ° 9	Ω	1	%	- (	0.005	%/V
1271	10 7	Ω	1	%	- (	0.005	%/V
1272	10 8 9	Ω	1	%	- (	0.005	%/V
1273	10 º 9	Ω	1	%	- (	0.02	%/V
1274	10 10 9	Ω	1	%	- (	0.02	%/V
1275	10 11 9	Ω	1	%	- (	0.02	%/V
1276	10 12 9	Ω	5	%	- (	0.02	%/V
1277	10 <sup>13</sup> 9	Ω	5	%	- (	0.04	%/V
1278	10 14 9	Ω	10	%	- (	0.04	%/V

## **DKD Calibration**

The calibration resistor model 1270 can be supplied with a DKD certificate (German calibration service). The documented measurement results and tolerances are captured with standards and measurement instruments that are subject to regular comparison to the national standards of the Federal Rep. of Germany. The verification by the appointed state authorities is shown in the certificate itself as well as the calibration sign which is placed on the device.


Model 12DKD-1270

## **WKS Calibration**

The manufacturer test certificate (WKS) includes the proof of traceability for national standards as well as protocolling of measurement results and uncertainties. Model 12WKS-1270

# **Application**

# Guard ring electrode



### **Guard Circuit**

The guard connection is exemplified by a guard ring electrode.

Depending on the connection wiring the RESISTOMAT® 24508 makes it possible to determinate the surface or volume resistance of the test sample.

For the determination of the surface resistance the measuring electrode ① is connected to the "X" input, the guard ring ② is connected with the "U" input and the basic electrode 3 is connected with the guard input.

For the determination of the volume resistance the measuring electrode ① is connected with the "X" input, the guard ring ② with the guard input and the basic electrode 3 is connected with the "U" input.