

Fully Integrated Thermoplastic Genosensor for the Highly Sensitive Detection and Identification of Multi-Drug-Resistant Tuberculosis**

Hong Wang, Hui-Wen Chen, Mateusz L. Hupert, Pin-Chuan Chen, Proyag Datta, Tana L. Pittman, Jost Goettert, Michael C. Murphy, Diana Williams, Francis Barany, and Steven A. Soper*

Infectious diseases are a major global health burden accounting for approximately 15 million deaths annually, many from drug-resistant pathogenic agents, with a significant number of cases occurring in developing countries.^[1–7] In particular, the resurgence of tuberculosis (TB) has been accompanied by the rapid spread of multi-drug resistance TB (MDR-TB) resulting from *Mycobacterium tuberculosis* (Mtb) strains that fail to respond to the first-line drugs, rifampin and isoniazid. Currently, less than 5 % of around 0.5 million MDR-TB cases estimated globally are appropriately diagnosed and treated because in part of the long assay turnaround time associated with conventional culture-based drug susceptibility testing.^[8]

Quantitative polymerase chain reaction (qPCR), line probe assays, or home-brewed nucleic acid amplification tests (NAATs), have recently been used to identify MDR-TB. These tests provide shorter assay turnaround times compared to culture, but depend on sophisticated laboratory infrastructure and well-trained personnel to ensure accurate, reliable, and reproducible results. The world health organization (WHO) expert group recently recommended two NAAT line probe assays, *INNO-LiPA RifTB* from Innogenetics and *MTBDRplus* from Hain Lifescience, both of which employ multiplexed PCR reverse hybridization approaches. Line probe assays, however, are not designed to interrogate

a mixed population of drug-resistant and susceptible bacterial populations,^[9–11] because sequence-specific array hybridization is unable to detect low abundance single-base variations because of cross-hybridization artifacts, especially in high guanine-cytosine (GC) content regions.

Several groups have developed partial or fully integrated microfluidic devices for carrying out NAATs for infectious diseases.^[12–17] For example, Cepheid's GeneXpert MDR-TB system performs qPCR using Taqman probes for Mtb and five rifampin-resistance mutations, providing results in nearly 2 h.^[18] However, many of these devices are made from silicon/glass materials and thus require direct photolithographic processing to manufacture the desired structures, which increases the production cost of devices.

Herein, we describe a modular design approach for an assay and hardware to detect/identify MDR-TB. The molecular assay was designed to interrogate single-base variations in codons 516, 526, and 531 (using the numbering system of *E. coli rpoB*) in the rifampin-resistance determining region (RRDR) of the *rpoB* gene, which was used as surrogate markers for MDR-TB.^[19,20] The multi-step assay was carried out within a modular thermoplastic fluidic cartridge operated by the accompanying support peripherals packaged into a small instrument (1' × 1' × 1'). No operator intervention was required once the clinical sample (sputum) was loaded into the fluidic cartridge. This system offered advantages compared to existing NAAT systems. For example, to accommodate low-resource setting scenarios where operator expertise is limited, the hardware provided full process automation. The fluidic system consisted of a cartridge made from thermoplastics with the desired structures fabricated in a high-production format using molding to keep chip cost low for one-time use applications (in vitro diagnostics) appropriate for resource limited settings.

Additional design concepts employed in the fluidic cartridge included: 1) A hybrid modular architecture, which combined several task-specific modules interconnected to a fluidic motherboard with the material selected to optimize performance. 2) Most of the active elements were poised off-chip to keep the chip costs low. 3) The embossing step was used to not only create the fluidic network, but other necessary device components as well, such as the DNA extraction bed, thermal isolation grooves, valve seats, and waveguides. The fluidic cartridge could easily be reconfigured without requiring re-engineering to accommodate alternative assays. For example, a universal array module could be replaced with a colorimetric readout module to simplify the

[*] Dr. H. Wang, Dr. H.-W. Chen, Dr. P.-C. Chen, Prof. M. C. Murphy
Department of Chemistry and Mechanical Engineering
Louisiana State University (USA)

Dr. M. L. Hupert, Prof. S. A. Soper
Department of Biomedical Engineering and Chemistry
University of North Carolina, Chapel Hill
Chapel Hill, NC 27599 (USA)
E-mail: ssoper@unc.edu

Dr. P. Datta, Dr. J. Goettert
CAMD, Louisiana State University (USA)

T. L. Pittman, Prof. D. Williams
HRSA/BPHC/National Hansen's Disease Center, Laboratory
Research Branch @ Louisiana State University-SVM (USA)

Prof. F. Barany
Department of Microbiology, Weill Cornell Medical College (USA)

[**] The authors would like to thank the NIH (EB010087) for partial financial support of this work as well as the World Class University (WCU) program of South Korea. Finally, the authors would like to thank Don Patterson, Tana Pittman, Weislaw Stryjewski, and Brad Ellison for this technical assistance.

Supporting information for this article is available on the WWW under <http://dx.doi.org/10.1002/ange.201200732>.

readout hardware. Compared to monolithic^[12–16] or LEGO approaches,^[21,22] the hybrid modular architecture balanced the efficiency and flexibility in fluidic cartridge design and material selection.

Collectively, more than 95 % of resistance to rifampin is associated with missense, insertion, and deletion mutations in the RRDR of the *rpoB* gene. Single-base variations in codons 516, 526, or 531 are most frequently found in rifampin-resistant *Mtb* strains world-wide. We therefore developed a PCR/ligase detection reaction (LDR)/universal array assay to detect sequence variations harbored within these codons (see Figure 1).^[23]




Figure 1. The molecular assay. Following a primary PCR to generate a 193 bp *rpoB* amplicon spanning the RRDR, three sets of ligase detection reaction (LDR) primers were used for interrogating missense mutations in codons 516, 526, and 531 of the *rpoB* gene. The discriminating primer contained a base complementary to either the wild-type or mutation sequence at its 3'-end. The common primer was phosphorylated at its 5'-end and contained a fluorescent dye, Cy5, at its 3'-end. A thermally stable ligase covalently links the two primers hybridized to the target if there was a perfect match at the locus being interrogated. The discriminating primer also carried a unique complementary "zipcode" sequence (cZip) at its 5'-end to direct the LDR product to a specific location on the "zipcode" universal array.

The PCR/LDR/universal array approach employed a high fidelity *Taq* ligase and decoupled the mutation discrimination step from the amplification and hybridization steps.^[23–27] The assay demonstrated the following important attributes: 1) The closely clustered drug resistance mutations of MDR-TB could be interrogated using a uniplex PCR. Therefore, careful design of primers with similar melting temperatures (T_m s) and problems with uneven amplification produced by different targets was negated. 2) Only the specific drug-resistant sequence variations generate positive results. Silent mutations, which do not confer drug resistance, did not generate false positive results. 3) Hetero-resistant *Mtb* containing a low abundance of drug-resistant strains (around 1 %) in patients with emerging drug resistance could be identified because of

the high discriminatory capability of the ligase enzyme. 4) Two amplification steps were used, an exponential amplification associated with the PCR and a linear amplification resulting from the LDR, which improved the limit of detection of the assay.

The fluidic cartridge shown in Figure 2, S1, and S2 (see the Supporting Information) was designed to include five different processing steps; cell lysis, solid-phase DNA extraction




Figure 2. Integrated, modular fluidic cartridge for TB analysis. A) 3D rendering of the cartridge and the array. 1–7: fluidic inlets and outlets: 1 = sample inlet, 2 = PCR cocktail inlet, 3 = LDR cocktail inlet, 4 = ethanol and air inlet, 5 = array wash inlet, 6 = vacuum connection, and 7 = waste. V_1 – V_6 = on-chip membrane valves (note that V_2 is positioned next to the SPE module on the cell lysis microchannel and is not visible in the current view). B) Close-up of the SPE bed showing the DNA capture bed filled with an array of high-aspect ratio pillars. C) Schematic operation of the on-chip membrane valve with mechanical actuation—electrically actuated solenoid presses on the polymer membrane closing the passage of fluid from the bottom layer through the valve and back to the bottom layer. D) Geometry of the continuous flow PCR reactor with dual-depth microchannels for extended residence time in the extension-zone; Den = denaturation, Ext = extension, and PA = primer annealing. E) Schematic representation of the detection mode. The laser excitation is coupled to a waveguide through an integrated prism. Light travelling through the waveguide excites the labeled LDR products hybridized to the zip code array spotted on the waveguide surface.

(SPE), PCR, LDR, and universal array hybridization. The thermal steps (cell lysis, PCR, and LDR) were situated on a fluidic motherboard while the SPE and universal array were placed on two different modules interconnected to the motherboard. A detailed description of how each module and the motherboard were fabricated is included in the Supporting Information. The modules and motherboard were hot embossed from a particular thermoplastic, selected to optimize performance of each specific processing step. To minimize post-assembly steps, embossing was used not only to generate fluidic networks, but also specific components required for proper operation of the module or motherboard. For example, the SPE module required the use of polycarbonate, which was suitable for the selective capture of nucleic acids to its surface following photoactivation.^[28]

Further, to increase the available surface area required to increase the DNA load, high-aspect ratio micropillars were manufactured into the SPE bed during embossing of the fluidic network, thus obviating the need for packing the SPE bed with externally added beads (see Figure S1 in the Supporting Information).

Polycarbonate has a relatively high glass transition temperature, which makes it a suitable material for thermal reactions and thus, the ideal material for the fluidic motherboard, which performed thermal reactions requiring operation temperatures of 65 to 95 °C. Polycarbonate also has a relatively large elongation at break threshold such that it can be used as a microfluidic valve membrane. Using polycarbonate as the cover plate for the motherboard obviated the need for an additional post-molding assembly step as the valve membranes were attached to the fluidic network in the same lamination step used to enclose the fluidic network with the cover plate. However, polycarbonate is not compatible with ultrasensitive fluorescence detection because of its relatively high background.

Poly(methyl methacrylate), PMMA, shows good optical clarity and minimal nonspecific adsorption artifacts, making it an ideal material for construction of the universal array module.^[29] The universal array was poised within a micro-channel resident on this module to reduce the time for probe addressing by minimizing diffusional constraints.^[30–32] Waveguides were fabricated using double-sided embossing with simple plasma activation of the PMMA surface to allow for covalent attachment of the zipcode probes to the waveguide (Figure S3 in the Supporting Information).^[33] Finally, printing of the zipcode probes could be performed using conventional DNA spotting equipment prior to enclosure of the fluidic network because of the low temperature required for bonding the cover plate to the substrate containing the probes.^[34]

Figure 2B and Figure S1B provide schematic drawings and SEM images of the SPE module. Design specifics on the SPE module can be found in the Supporting Information. Figure 2C shows a schematic of the operation of the polycarbonate membrane valves with solenoid actuation. These microvalves could withstand head pressures up to 10⁵ psi without leakage, which is a more than an order of magnitude higher pressure load relative to that of polydimethylsiloxane (PDMS) valves.^[35]

The thermal reactors situated on the fluidic motherboard (Figure 2A, and Figures S1A and S1C in the Supporting Information) incorporated a continuous flow format, which provided ultrafast PCR amplification because extension times are limited by the kinetics of the polymerase for properly designed thermal reactors.^[36] Several thermal management structures were also included to further improve amplification efficiency (Figure S4 in the Supporting information), such as backside thermal isolation grooves, a thin substrate, and copper plates to give a uniform temperature distribution throughout a particular thermal zone.^[37] Further, a dual-depth channel (200 and 100 µm) was employed to provide sufficient residence time within the polymerase extension zone to generate full-length PCR products during each cycle while operating at a fixed linear flow velocity (Figure 2D and Figure S1C in the Supporting information). For a detailed

discussion on the operation of the fluidic cartridge, see the Supporting Information.

We performed a series of assays using different positive and negative controls that were processed using the modular fluidic cartridge. The results concurred with those expected, as shown in Figure S5 in the Supporting Information. For example, supplying an input of *E. coli* cells did not result in positive fluorescence signals on the universal array. However, input of a wild-type *Mtb* strain generated fluorescence signatures at the appropriate spots of the array (Figure S5 in the Supporting Information).

We were interested in evaluating the minimum number of *Mtb* cells required to provide positive results using our system and whether results could meet the needed level for clinical diagnosis. To determine the limit-of-detection, assays were conducted using a serial dilution of cultured H37Rv cells as the input. As few as 50 *Mtb* cells could be successfully detected, representing a 100-fold improvement in sensitivity compared to current clinical smear tests, which require 5000–10000 bacilli in 1 mL of sputum.

Hetero-resistant *Mtb* is defined as the coexistence of mixed populations of drug-susceptible and resistant *Mtb* strains in the same patient, which is considered as a preliminary stage to full resistance. These cases result either from super-infection or from selection pressure during antibiotic treatment. The occurrence of hetero-resistance can be as high as 20%.^[10,11] To evaluate the ability of our system to identify a minority population of drug-resistant *Mtb* strains, a series of control hetero-resistance samples containing increasing percentages of a rifampin-resistant strain (S531L)—0, 1, 2, 5, 10, and 100%—in a drug-susceptible strain of H37Rv were prepared and analyzed (Figure 3). The fluorescence intensity from the 531 mutant spot increased with an increased percentage of the rifampin-resistant strain present in the sample while the spot intensity from the 531Wt decreased (Figure 3C). The fluorescence from the 531 mutant spots run with samples containing as little as 1% of rifampin-resistant strain was calculated to be around 10-fold higher than the fluorescence from those run with samples containing 0% rifampin-resistant strains (Figure 3C). Therefore, less than 1% of the drug-resistant strain could still be discriminated based on a three-fold standard-deviation higher signal than a nonspecific signal. Very slight levels of fluorescence were detected from the 531Wt spots run with 100% of the drug-resistant strain, which could be attributed to misligation and/or nonspecific hybridization.^[38]

The modular system was challenged with six clinical samples containing either rifampin-resistant or susceptible *Mtb* strains. Figure 4 represents results that were consistent with the phenotype and the genotype of those clinical samples. The analysis of samples YE69 (A516V), 3976-83 (A516V), YE68 (H526Y), 3908-83 (S531L), and 94-2219 (Wt), Figure 4B–G, identified the correct mutation for each drug-resistant strain as determined by sequence analysis of the *rpoB* gene of these clinical isolates.

A colorimetric zipcode array module was developed as detailed in the Supporting Information to simplify the readout by eliminating the need for laser-induced fluorescence (Figure S6 in the Supporting Information). PMMA




Figure 3. Identification of a mixed-population of drug-susceptible (H37Rv) and drug-resistant (S531L) *Mtb* strains. A) Layout of the universal array. Spots labeled C were printed with 5'-Cy5 oligonucleotides, and used as quality controls (C) of the printing process. Spots labeled as Hyb were hybridization controls. Spots labeled as Pos were positive controls (spiked plasmids). Spots labeled Mtb were *Mtb*-specific probes targeting an IS6110 insertion fragment. B–G) Universal array images for different amounts of the drug-resistant and drug-susceptible strains. The percentage of the sample containing the drug resistant strain was 0 (B), 1 (C), 2 (D), 5 (E), 10 (F), and 100% (G).

wafers containing immobilized zipcode probes were prepared using a PDMS stencil (Figure S6B in the Supporting Information). Also, 1.4 nm nanogold-labeled common primers were used instead of fluorescent dye-labeled primers as shown in Figure S6A in the Supporting Information. After LDR and hybridization, the nanogold labels acted as a catalytic site for silver deposition, which was visible to the naked eye and could be recorded by a digital camera. A drug-susceptible *Mtb* stain (H37Rv) and a rifampin-resistant strain (S531L) were tested and the results are shown in Figure S6C in the Supporting Information. Using this colorimetric test for detection, the *Mtb* cell limit-of-detection was found to be 50 cells, similar to that observed using laser-induced fluorescence detection.

We have reported the development of a modular design approach for both the software (molecular assay) and hardware (fluidic cartridge and support peripherals) to determine sequence variations in reporter sequences to detect and identify MDR-TB directly from sputum samples. The system provided full-process automation minimizing operator expertise requirements and also, generated results in less than 30 min, which will be a key operating metric by providing rapid results to allow proper treatment of patients, potentially minimizing the generation of additional drug-resistant strains because of the implementation of a full treatment regimen.




Figure 4. Universal array hybridization results from clinical samples (sputum). A) Layout of the universal array with the same designations as that given in Figure 3A. B) *Mtb*(+) clinical isolate (YE69) harbors a mutation in codon 516 of *rpoB*. C) *Mtb*(+) sputum sediment (3976-83) harbors a mutation in codon 516 of *rpoB*. D) *Mtb* clinical isolate (YE68) possesses a mutation in codon 526 of *rpoB*. E) *Mtb*(+) sputum sediment (3908-83) harbors a mutation in codon 531 of *rpoB*. F) *Mtb*(+) sputum sediment (94-2219) contained no mutations in *rpoB* RRDR (Wt). G) *Mtb*(-) sputum sediment.

Recently, extremely drug-resistant TB (XDR-TB) has been reported. XDR-TB strains contain mutations responsible for MDR-TB as well as sequence variations resulting in resistance to second-line drugs. By incorporating a multiplexed PCR and additional LDR primers, the system presented here can easily be configured to detect sequence variations responsible for XDR-TB without hardware redesign. The system, both software and hardware, holds the potential to be a universal platform for identifying genetic sequence signatures in a variety of applications, including cancer diagnostics,^[23,26,27] forensic testing, and biothreat pathogen detection, and identification in both developing and developed countries^[24] because of the modular design.

Received: January 26, 2012

Published online: March 19, 2012

Keywords: genotyping · fluorescence · microfluidics · tuberculosis

- [1] World Health Organization, "Report on infectious diseases: Removing Obstacles to Healthy Development", can be found under <http://www.who.int/infectious-disease-report/index-rpt99.html>, 1999.
- [2] World Health Organization, "The world health report 2002—Reducing Risks, Promoting Healthy Life", can be found under <http://www.who.int/whr/2002/en/>, 2002.

- [3] World Health Organization, "Scaling up the response to infectious diseases: A way out of poverty", can be found under <http://www.who.int/infectious-disease-report/2002/>, **2002**.
- [4] M. D. Perkins, G. Roscigno, A. Zumla, *Lancet* **2006**, *367*, 942–943.
- [5] D. M. Morens, G. K. Folkers, A. S. Fauci, *Nature* **2004**, *430*, 242–249.
- [6] I. N. Okeke, K. P. Klugman, Z. A. Bhutta, A. G. Duse, P. Jenkins, T. F. O'Brien, A. Pablos-Mendez, R. Laxminarayan, *Lancet Infect. Dis.* **2005**, *5*, 568–580.
- [7] I. N. Okeke, R. Laxminarayan, Z. A. Bhutta, A. G. Duse, P. Jenkins, T. F. O'Brien, A. Pablos-Mendez, K. P. Klugman, *Lancet Infect. Dis.* **2005**, *5*, 481–493.
- [8] World Health Organization, "Policy guidance on drug-susceptibility testing (DST) of second-line antituberculosis drugs", can be found at http://www.who.int/tb/features_archive/xdr_mdr_policy_guidance/en/index.html, **2008**.
- [9] World Health Organization, "Expert Group Report. Molecular Line Probe Assays for Rapid Screening of Patients at Risk of Multi-Drug Resistant Tuberculosis (MDR-TB)", can be found at http://www.who.int/tb/features_archive/expert_group_report_june08.pdf, **2008**.
- [10] S. Hofmann-Thiel, J. van Ingen, K. Feldmann, L. Turaev, G. T. Uzakova, G. Murmusaeve, D. van Soolingen, H. Hoffmann, *Eur. Respir. J.* **2009**, *33*, 368–374.
- [11] H. Rinder, K. T. Mieskes, T. Loscher, *Int. J. Tuberc. Lung Dis.* **2001**, *5*, 339–345.
- [12] A. F. Sauer-Budge, P. Mirer, A. Chatterjee, C. M. Klapperich, D. Chargin, A. Sharon, *Lab Chip* **2009**, *9*, 2803–2810.
- [13] C. J. Easley, J. M. Karlinsey, J. M. Bienvenue, L. A. Legendre, M. G. Roper, S. H. Feldman, M. A. Hughes, E. L. Hewlett, T. J. Merkel, J. P. Ferrance, J. P. Landers, *Proc. Natl. Acad. Sci. USA* **2006**, *103*, 19272–19277.
- [14] R. H. Liu, J. N. Yang, R. Lenigk, J. Bonanno, P. Grodzinski, *Anal. Chem.* **2004**, *76*, 1824–1831.
- [15] R. Pal, M. Yang, R. Lin, B. N. Johnson, N. Srivastava, S. Z. Razzacki, K. J. Chomistek, D. C. Heldsinger, R. M. Haque, V. M. Ugaz, P. K. Thwar, Z. Chen, K. Alfano, M. B. Yim, M. Krishnan, A. O. Fuller, R. G. Larson, D. T. Burke, M. A. Burns, *Lab Chip* **2005**, *5*, 1024–1032.
- [16] R. G. Blazej, P. Kumaresan, R. A. Mathies, *Proc. Natl. Acad. Sci. USA* **2006**, *103*, 7240–7245.
- [17] P. Yager, T. Edwards, E. Fu, K. Helton, K. Nelson, M. R. Tam, B. H. Weigl, *Nature* **2006**, *442*, 412–418.
- [18] P. Ioannidis, D. Papaventsis, S. Karabela, S. Nikolaou, M. Panagi, E. Raftopoulou, E. Konstantinidou, I. Marinou, S. Kanavaki, *J. Clinical Microbiol.* **2011**, *49*, 3068–3070.
- [19] D. L. Williams, L. Spring, L. Collins, L. P. Miller, L. B. Heifets, P. R. J. Gangadharam, T. P. Gillis, *Antimicrob. Agents Chemother.* **1998**, *42*, 1853–1857.
- [20] D. L. Williams, C. Waguespack, K. Eisenach, J. T. Crawford, F. Portaels, M. Salfinger, C. M. Nolan, C. Abe, V. Stichtgroh, T. P. Gillis, *Antimicrob. Agents Chemother.* **1994**, *38*, 2380–2386.
- [21] M. Rhee, M. A. Burns, *Lab Chip* **2008**, *8*, 1365–1373.
- [22] K. A. Shaikh, K. S. Ryu, E. D. Goluch, J. M. Nam, J. W. Liu, S. Thaxton, T. N. Chiesl, A. E. Barron, Y. Lu, C. A. Mirkin, C. Liu, *Proc. Natl. Acad. Sci. USA* **2005**, *102*, 9745–9750.
- [23] N. P. Gerry, N. E. Witowski, J. Day, R. P. Hammer, G. Barany, F. Barany, *J. Mol. Biol.* **1999**, *292*, 251–262.
- [24] M. Pingle, M. Rundell, S. Das, L. M. Golightly, F. Barany, *Methods Mol. Biol.* **2010**, *632*, 147–157.
- [25] S. Das, M. R. Pingle, J. Muñoz-Jordán, M. S. Rundell, S. Rondini, K. Granger, G. J. Chang, E. Kelly, E. G. Spier, D. Larone, E. Spitzer, F. Barany, L. M. Golightly, *J. Clin. Microbiol.* **2008**, *46*, 3276–3278.
- [26] Y. W. Cheng, C. Shawber, D. Notterman, P. Paty, F. Barany, *Genome Res.* **2006**, *16*, 289–292.
- [27] R. Favis, J. P. Day, N. P. Gerry, C. Phelan, S. Narod, F. Barany, *Nat. Biotechnol.* **2000**, *18*, 561–564.
- [28] M. A. Witek, S. D. Llopis, A. Wheatley, R. L. McCarley, S. A. Soper, *Nucleic Acids Res.* **2006**, *34*, e74.
- [29] F. Xu, P. Datta, H. Wang, S. Gurung, M. Hashimoto, S. Wei, J. Goettert, R. L. McCarley, S. A. Soper, *Anal. Chem.* **2007**, *79*, 9007–9013.
- [30] Y. Wang, B. Vaidya, H. D. Farquar, W. Stryjewski, R. P. Hammer, R. L. McCarley, S. A. Soper, Y. W. Cheng, F. Barany, *Anal. Chem.* **2003**, *75*, 1130–1140.
- [31] C. Situma, Y. Wang, M. Hupert, S. A. Soper, *Abstr. Pap. Am. Chem. Soc.* **2004**, *228*, U118–U118.
- [32] C. Situma, Y. Wang, M. Hupert, F. Barany, R. L. McCarley, S. A. Soper, *Anal. Biochem.* **2005**, *340*, 123–135.
- [33] R. L. McCarley, B. Vaidya, S. Y. Wei, A. F. Smith, A. B. Patel, J. Feng, M. C. Murphy, S. A. Soper, *J. Am. Chem. Soc.* **2005**, *127*, 842–843.
- [34] S. M. Ford, J. Davies, B. Kar, S. D. Qi, S. McWhorter, S. A. Soper, C. K. Malek, *J. Biomech. Eng.* **1999**, *121*, 13–21.
- [35] W. Zhang, S. Lin, C. Wang, J. Hu, C. Li, Z. Zhuang, Y. Zhou, R. A. Mathies, C. J. Yang, *Lab Chip* **2009**, *9*, 3088–3094.
- [36] M. Hashimoto, P. C. Chen, M. W. Mitchell, D. E. Nikitopoulos, S. A. Soper, M. C. Murphy, *Lab Chip* **2004**, *4*, 638–645.
- [37] P. C. Chen, D. E. Nikitopoulos, S. A. Soper, M. C. Murphy, *Biomed. Microdevices* **2008**, *10*, 141–152.
- [38] J. Y. Luo, D. E. Bergstrom, F. Barany, *Nucleic Acids Res.* **1996**, *24*, 3071–3078.