

A One-Drop Cryoscope: The Tonicity of Frog and Goldfish Sera<sup>1</sup>

A very promising device which provides objective register of freezing point depressions *via* recording potentiometer on relatively small samples (0.05–0.1 ml) and an automatic nucleation feature to prevent under-cooling, has recently been described<sup>2</sup>. This instrument, however, was not used with biological materials. This paper will describe its use with such materials and present a simple modification in the design of the sample chamber for samples of 0.02 ml.

The modified sample chamber was essentially a glass tube with a thermistor sealed in its bottom. A glass tube (4 mm I.D. by 15 mm long) was one-third filled with liquid silicone rubber sealant (Dow Silastic 891) and allowed to harden. The thermistor probe (Glennite 32PB2) was prepared by grinding off most of its glass covering with a motorized fine stone. The thermistor leads were threaded through the bore of a short 18-gauge needle which was forced through the hardened sealant. The needle was removed and the probe pulled through the silicone rubber by its leads until the thermistor bead just reached the top of the sealant. A rubber grommet adapted the chamber to a larger diameter (7 mm I.D.) tube to provide a holder. Finally, leads and sample chamber were sealed in their holder with more sealant.

The design of the 20-μl chamber was shown to be valid by demonstrating that linearity obtained between concentration (in the biological range) and scale readings of the freezing points. A 100 mOsm concentration difference equalled 30 scale units on the recorder<sup>3</sup> chart. It was further demonstrated that published values for the tonicity of a representative organism, *Rana pipiens*, may be reproduced with ease by use of the modified chamber.

Frogs, *Rana pipiens*, in apparent good health were obtained from a local supplier and kept in a laboratory tank containing a small amount of tap water (equivalent to

20 mOsm KCl) for four days at 23°C before bleeding. Blood was drawn directly from the ventricles and centrifuged under oil to pack the clot and cells and remove other potential nucleation centers. 20-μl samples of the serum were run in the cryoscope and the results are presented in the Table. Frog serum has an average tonicity equivalent to 200 mOsm KCl or to 0.65 ± 0.05% NaCl. This value agrees with that obtained by KROGH<sup>4</sup> and MACALLAM<sup>5</sup>. It is lower than that reported by HOBER<sup>6</sup> and ADOLPH<sup>7</sup>. However, the former was derived from frogs held in a more concentrated environment than ours (equivalent to 57 mOsm KCl) whereas the latter was an average synthesized from a number of species.

The equivalent tonicity for the serum of the goldfish *Crassius auratus* (L.), standard xanthic, common comet was also determined. The value obtained (Table) was equivalent to 289 mOsm KCl or to 0.92 ± 0.05% NaCl. Although this value has not been previously reported, its correctness is substantiated by the following. The measurement falls in the general range of that reported for fresh-water teleosts<sup>8</sup>, and luxurious *in vitro* growth of goldfish tails occurs in the media used for human skin<sup>9</sup>, whose tonicity falls in the 0.9% NaCl range<sup>10</sup>.

*Zusammenfassung.* Es wird eine einfache Modifikation eines Kryoskopes zur objektiven Messung von Gefrierpunktserniedrigungen im Bereiche von 20 μl beschrieben. Das Instrument eignet sich u.a. zur Bestimmung der Gefrierpunktserniedrigungen des Wirbeltierblutes.

L. LEVINE and DIANA A. MUSALLAM

Department of Biology, Wayne State University, Detroit (Michigan USA) and Friends Boys' School, Ramallah (Jordan), January 20, 1964.

The tonicity of frog and goldfish sera, in equivalent osmolality and % NaCl

Animal	Number of determinations	Average Equivalent	
		KCl (mOsm) ± standard error	NaCl % ± confidence interval (95%)
<i>Rana pipiens</i>	28	200 ± 4.8	0.65 ± 0.05
<i>Crassius auratus</i>	11	289 ± 4.3	0.92 ± 0.05

Rice Blast, Nyctotemperature, and Nitrate Reductase

That a particular genotype-nyctotemperature combination is an essential prerequisite for the occurrence of 'blast' disease of rice (*Oryza sativa* L.) caused by *Piricularia oryzae* Cav. has already been shown in this laboratory<sup>1</sup>. Subjecting susceptible rice plants (CO 13) to a

nyctotemperature regime of 20°C gave the necessary pre-disposition for successful infection by the 'blast' fungus and in such plants there was an accelerated tempo of nitrogen metabolism with a considerable synthesis of the amide glutamine<sup>2</sup>. Conversely, when susceptible plants

<sup>1</sup> S. SURYANARAYANAN, Proc. nat. Inst. Sci. India 24 B, 285 (1958).  
<sup>2</sup> S. SURYANARAYANAN, Curt. Sci. 27, 447 (1958).