

# Abstracts

## A Dual Six-Port Automatic Network Analyzer (1980 [MWSYM])

---

*H.M. Cronson and L. Susman. "A Dual Six-Port Automatic Network Analyzer (1980 [MWSYM])." 1980 MTT-S International Microwave Symposium Digest 80.1 (1980 [MWSYM]): 434-436.*

A 2 to 18 GHz dual six-port automatic network analyzer is described. The measurements showed accuracies better than 0.1 dB up to 40 dB insertion loss from 2 to 7 GHz. The cause of deviations at higher frequencies is discussed.

[Return to main document.](#)