

# Abstracts

## Non-Pertubing Temperature Probe and Thermography Measurements in Microwave Diathermy

---

*C.U. Hochuli and G. Kantor. "Non-Pertubing Temperature Probe and Thermography Measurements in Microwave Diathermy." 1981 MTT-S International Microwave Symposium Digest 81.1 (1981 [MWSYM]): 470-472.*

In support of the BRH proposed microwave diathermy standard, temperature measurements in fat-muscle phantoms using thermography and a non-perturbing temperature probe were investigated. These probe data indicate significant thermography errors.

[Return to main document.](#)