

Abstracts

Measurement of Dielectric Properties of Biological Substances Using Improved Open-Ended Coaxial Line Resonator Method

X. Dening, L. Liping and J. Zhiyau. "Measurement of Dielectric Properties of Biological Substances Using Improved Open-Ended Coaxial Line Resonator Method." 1987 MTT-S International Microwave Symposium Digest 87.1 (1987 Vol. I [MWSYM]): 251-254.

An improved open-ended coaxial line resonator method is presented. By considering the influences of radiation effect and higher order modes, the measurement frequency range has successfully extended to X-band. The results of seven organs of a dog have given, which can be referred in bioelectromagnetics.

[Return to main document.](#)