

# Abstracts

## Development of Broadband Coaxial Components for a 0.04 to 110 GHz Measurement System

---

*B. Oldfield. "Development of Broadband Coaxial Components for a 0.04 to 110 GHz Measurement System." 1995 MTT-S International Microwave Symposium Digest 95.3 (1995 Vol. III [MWSYM]): 1471-1473.*

Coaxial components including a new connector system (the W connector), couplers, switches, terminations and Waveguide to coax adapters, were used to configure a VNA with continuous coverage from 0.04 to 110 GHz. Wafer probes covering the same range are also described.

[Return to main document.](#)