

Fig. 1—Radial line cavity assembly.

were constructed of brass to the configuration of Fig. 1. Center-of-band rejection was at least 25 db in all cases; this was adequate for our purposes. For greater rejection, electro-formed cavities are described by de Loach should be adopted, and to broaden the rejection band a number of cavities can be coupled in series.

The author is indebted to the Managing Director of G. and E. Bradley, Ltd. for permission to publish this communication. He also wishes to thank R. G. Amos and J. A. Dodson for making the necessary measurements.

M. A. R. GUNSTON
G. and E. Bradley, Ltd.
London, England

Correction to "Electrolytic Pointing of Fine Wire"

In the above correspondence,¹ the final line should have read 0.0002 inch instead of 0.002 inch.

J. W. DOZIER
Advanced Technology Corp.
Timonium, Md.

Manuscript received July 9, 1964.

¹ J. W. Dozier and J. D. Rodgers, IEEE TRANS. ON MICROWAVE THEORY AND TECHNIQUES, vol. MTT-12, p. 360; May, 1964.

Microwave and High-Frequency Calibration Services of the National Bureau of Standards—Part II

INTRODUCTION

Following the series of presentations on microwave and high-frequency calibration services of the National Bureau of Standards which began in the July, 1964 issue of these TRANSACTIONS, the services for the measure-




Fig. 1—Microwave attenuation calibrations (rectangular waveguide).




Fig. 2—High-frequency attenuation calibrations (coaxial).

