



# Supplement

Manual Title:	DSP-FTA420/30/40 Users	Supplement Issue:	1
Part Number:	1570398	Issue Date:	12/01
Print Date:	July 2000	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

# Change #1

On page 3,



Change:   **Warning**



To:   **Warning** CLASS 1 LASER PRODUCT

Add:

- Do not start any optical tests before completely connecting all of the fiber optic cables.

On page 42, under General Specifications,

Change:  

To:   Complies with EN60825-2 and FDA 21 CFR Subchapter J

Add:   CLASS 1 LASER PRODUCT

Under Optical Transmitter, Power output,

Change: FTA420 -20 dBm (10  $\mu$ W), typical (into 62.5/125  $\mu$ m fiber)

To: FTA420 -20 dBm typical (into 62.5/125  $\mu$ m fiber)

Change: FTA430 and FTA440 -10 dBm

To: FTA430 and FTA440 -8 dBm typical