

## S E R V I C E N O T E

SUPERSEDES: NONE

**3324A Synthesized Function/Sweep Generator 21 MHz****Serial Numbers:**

0000G00000 / 2836G00510

0000A00000 / 9999A99999

**Installation of Phase Calibration Option #003 or #004****Parts Required:****Part No.****Description**

03324-68713

Phase Calibration Slave (#003)

03324-68714

Phase Calibration Master (#004)

**Situation:**

Phase Calibration Option 003/004 are not field-installable in 3324As with serial numbers shown above. It is possible to have these instruments upgraded with option #003 or #004 at the factory.

**Solution/Action:**

3324As with serial numbers shown above have to be sent back to Boeblingen Instruments Division (BID) to have the Phase Calibration Option #003 or #004 installed. Please order the option retrofit kit, and send the option together with the instrument to BID. At BID we will install the option and upgrade the instrument.

*Continued*

DATE: 03 December 1992

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

**INFORMATION ONLY**

AUTHOR:

ENTITY:

ADDITIONAL INFORMATION:

PW

B100



Please contact the Division before sending the instruments. You will find the responsible person at the division by checking the latest Service Organization Chart (SupOrg) which is available on AgilentDesk Autoanswer:

To receive the Autoanswer info, please perform the following steps:

```
Intray>      send
Subject:      SupOrg
TO:           BIDSVC F
FROM:         your name
Text:         //
MESSAGE      mail
```