

## S E R V I C E N O T E

SUPERSEDES: NONE

**54123A Four Channel Test Set**

54121A-02B Serials: 0000A00000 / 9999Z99999

54123A-01 Serials: 0000A00000 / US99999999

54124A-01 Serials: 0000A00000 / US99999999

**TDR System Specification Change****Duplicate Service Notes:** 54121A-02B, 54123A-01, 54124A-01

The 54123A TDR System Flatness specification has changed from:

“< +5%, -3% to 1 ns from edge (measured at 18 GHz bandwidth in average display mode)” to the following:

“< +5%, -5% to 1 ns from edge (measured at 18 GHz bandwidth in average display mode)”

Please make this - 3% to - 5% specification change in your 54123A service manual on the following pages: 1-4, 3-23, 3-26, 3-27 and test record page 3-35.

New instruments and replacement TDR Step Generators, part number 5086-7471 and 5086-6471, now have this specification change. This specification change does not degrade measurement capability of the 54120 TDR system.

DATE: January 1996

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

**INFORMATION ONLY**

AUTHOR:

ENTITY:

ADDITIONAL INFORMATION:

RK

0840

