

S E R V I C E N O T E

SUPERSEDES: None

54502A Oscilloscope**Serial Numbers:** 0000A00000 / 2934A05900**Test Failure:**

During the input resistance performance test the fifty ohm resistor measures slightly out of specification on the high side, i.e., 50.6 or 50.7 ohms. (Usually not greater than 51 ohms).

Duplicate Service Notes:

54502A-06

54503A-05

54504A-02

54510A-01

To Be Performed By: Agilent-Qualified Personnel or Customer**Parts Required:****Part No.**

54503-69102

Description

Grounding Spring (2 required per attenuator)

Continued

DATE: 04 May 1994

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:			
MODIFICATION RECOMMENDED			
ACTION CATEGORY:	<input type="checkbox"/> IMMEDIATELY <input checked="" type="checkbox"/> ON SPECIFIED FAILURE <input type="checkbox"/> AGREEABLE TIME	STANDARDS: Labor 1.5 Hour	
LOCATION CATEGORY:	<input checked="" type="checkbox"/> CUSTOMER INSTALLABLE <input type="checkbox"/> ON-SITE <input checked="" type="checkbox"/> SERVICE CENTER	SERVICE INVENTORY: <input type="checkbox"/> RETURN <input type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT	USED PARTS: <input type="checkbox"/> RETURN <input checked="" type="checkbox"/> SCRAP <input type="checkbox"/> SEE TEXT
AVAILABILITY:	PRODUCT'S SUPPORT LIFE		AGILENT RESPONSIBLE UNTIL: 01 October 1996
AUTHOR: TG	ENTITY: 0800	ADDITIONAL INFORMATION:	



Situation:

Due to elastomer compression, instruments in the above serial number range may fail the Input Resistance Performance Test as called out in the Service Manual.

Solution/Action:


Upon specified failure, or if the Processor board must be removed for any reason, perform the following procedure to replace the blue elastomeric grounding gaskets with gold-plated grounding springs:

1. Remove the two screws securing each Attenuator Assembly to the A1 Processor board.
2. Carefully ease each of the Attenuator Assemblies out of the connector on the A1 board.
3. Using a suitable implement, gently ease the two retaining spring clips from each attenuator assembly. (See Figure 2)
4. Using a pair of tweezers, grip one of the connector pins and gently ease the substrate assembly (connector end first) from the attenuator housing. (See Figure 2)
5. Inspect the substrate assembly for damage.
6. Remove the blue elastomeric gaskets (2 ea) from the gasket recesses in the base of the attenuator housing. (See Figure 2)
7. Using tweezers, fit the new gold-plated grounding springs into the gasket recesses.
8. Insert the substrate assembly into the housing, gold contact end first. Carefully locate the gold contact under the center conductor of the housing BNC connector. (See Figure 2)
9. Seat the substrate assembly in place and check that the gold-plated grounding springs are properly seated in the gasket recesses.

NOTE

To check for proper seating of the gold-plated grounding springs, slide the substrate assembly first to one side and then the other and check for the reflection of the grounding springs on the wall of the module housing.

10. Fit the spring clips back into place ensuring that they are sprung into position over the external side lugs. (See Figure 1)
11. When installing each module back onto the A1 board first ensure that the locating pins on the housing are fully inserted into the two holes on the board then insert and tighten the front screws.
12. Insert and tighten the rear screws.
13. Reassemble the instrument in accordance with the Service Manual.



54501E18

Attenuator Assembly
Figure 1




Figure 2