

S E R V I C E N O T E

SUPERSEDES: None

54520A Oscilloscope

Serial Numbers: 0000A00000/US99999999

During Self-Calibration the instrument fails Logic Trigger Delay Calibration.

Duplicate Service Notes: 54540A-04
 54520C-05
 54540C-05

Situation:

If any other channel is turned on in addition to channel one, the instrument will fail the Logic Trigger Delay Calibration.

Solution/Action:

Before performing a Self-Calibration, press the "Recall" and "Clear" softkeys. This will place the instrument in a default state with channel one on and all other channels off.

DATE: March 1997

ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

INFORMATION ONLY

AUTHOR:

TG

ENTITY:

0800

ADDITIONAL INFORMATION: