

## S E R V I C E N O T E

SUPERSEDES: None

**54520C Oscilloscope**

Serial Numbers: 0000A00000/US99999999

**During Self-Calibration the instrument fails Logic Trigger Delay Calibration.****Duplicate Service Notes:** 54540A-04  
54520A-04  
54540C-05**Situation:**

If any other channel is turned on in addition to channel one, the instrument will fail the Logic Trigger Delay Calibration.

**Solution/Action:**

Before performing a Self-Calibration, press the "Recall" and "Clear" softkeys. This will place the instrument in a default state with channel one on and all other channels off.

DATE: March 1997

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

**INFORMATION ONLY**

AUTHOR:

TG

ENTITY:

0800

ADDITIONAL INFORMATION:

