

## S E R V I C E N O T E

SUPERSEDES: None

**HP 54522C Oscilloscope****Serial Numbers:** 0000A00000 / US99999999**When Power is cycled during an HP-IB activity the Oscilloscope will lock up.****Duplicate Service Notes:**

54520C-04	54540A-03
54542A-03	54520A-03
54522A-03	54540C-04
54542C-04	

**Situation:**

If power is cycled while the instrument is performing operations over HP-IB, the instrument will lock-up.

**Solution/Action:**

Power the instrument down, and while holding down any front panel key, power the instrument back up. This will clear the lock-up problem and allow normal operation of the instrument.

DATE: March 1997

## ADMINISTRATIVE INFORMATION

SERVICE NOTE CLASSIFICATION:

**INFORMATION ONLY**

AUTHOR:

ENTITY:

ADDITIONAL INFORMATION:

TG

0800